
PRODUCT TECHNICAL DATASHEET

ITEM REF:	011-CH-33	REVISION:	01

icumedical
human connections

LEGAL MANUFACTURER	**ICU Medical, Inc.** 951 Calle Amanecer, San Clemente, CA 92673, USA
IMPORTER	**ICU Medical B.V.** Hofspoor 3, 3994 VZ Houten, The Netherlands
EU AUTHORISED REPRESENTATIVE (EUAR)	**Medical Device Safety Service GmbH (MDSS)**, Schiffgraben 41, 30175 Hannover, Germany
ASSEMBLY SITE	**ICU Medical de Mexico**, S.A. de C.**.,** Avenida Cuarza No. 250, Colonia Rancho Santa Clara, Delegacion Manaedero Ensenada, Baja California, Mexico 22790
CLASSIFICATION CODE	**GMDN** Code: 41222
INTENDED USE	Medication transfer set A collection of sterile devices and supplies designed to transfer parenteral medications [e.g., intravenous (IV) fluids, drugs] between a first container(s) [e.g., a vial(s)] and a second container/administration line [e.g., an IV bag/line] in a non-powered closed system; it may also be used to compound and/or reconstitute medication for its preparation and administration. It is available in a variety of configurations and typically includes vials, tubulures, connectors, spikes, syringes, and caps. This is a single-use device.
ITEM DESCRIPTION	Double Clave™ Syringe Transfer Set

PRIMING VOLUME (ml)	0.15	**LENGTH (cm)**	5.81	**WEIGHT (g)**	3.86	**CASE QTY**	50



ITEM SPECIFIC DATA	**MRI Compatibility**	No metal components
	Chemical Compatibility	Lipids & Common Chemotherapeutics
	Luer Compatibility	ISO 80369-7 Compliant.
	Sterilization and Shelf Life	Radiation; 5-Year Expiration
CLAVE SPECIFIC DATA	**Backpressure Rating**	45 psig / 2327 mmHG (Unactivated)
	Disinfection Compatibility	70% Isopropyl Alcohol
	Luer Compatibility	ISO 80369-7 Compliant male luers > 1,55mm Internal diameter

PRODUCT TECHNICAL DATASHEET			
ITEM REF:	**011-CH-33**	**REVISION:**	**01**

icumedical
human connections

LIST OF COMPONENTS *(Latex and DEHP Free)*	1	CONNECTOR, DOUBLE, FEMALE/FEMALE	POLYCARBONATE
	2	MICROCLAVE®, BLUE	POLYCARBONATE SILICONE SILICONE LUBRICANT
MATERIAL COMPLIANCE AND BIOCOMPATIBILITY	Product has met the requirements of ISO 10993-1 and related standards for biocompatibility.		
PRECAUTIONS	Use aseptic techniques. Single-use only – Do not resterilize. Sterile, Non-Pyrogenic fluid pathway in unopened, undamaged package.		
LABELS AND DIRECTIONS FOR USE	Label and/or DFU contain information for proper use including any warnings or contraindications. Labels are applied on the individual package and on the outside of the cardboard box. Each sales package contains a Direction for Use where applicable.		
PACKING AND PACKAGING	The non-unit packaging is biodegradable as defined by Directive 94/62/EC. Packaging met the requirements of BS EN ISO 11607. The packaging for this product is Latex Free. This item is blister packed or pouched individually using medical grade materials.		
TRACEABILITY	Lot number provides full traceability of all components and manufacturing processes.		
STORAGE	Store in a dry and clean place. Product should be retained in packaging until ready for use.		
DISPOSAL	The user must dispose of the device according to hospital disposal policy.		
PRODUCTION AND ENVIRONMENT CONTROLS	• Product is manufactured in an environmentally controlled room. Floor, surfaces, and environment are monitored according to quality procedures. • Each component is inspected during acceptance (example: visual; dimensional; functional) to verify conformity with the requirements according to quality procedures. • Production and release specific tests are performed according to quality procedures. • Limulus amebocyte lysate (LAL) testing is performed on production samples to verify conformity of the endotoxin charge according to quality procedures. • Bioburden tests are performed on production samples to verify conformity of the microbial charge according to quality procedures.		
QUALITY SYSTEM AND PRODUCT CERTIFICATION	**Quality System complies to:**	ISO 13485:2016	
	Product Certification:	The product is manufactured in compliance to Council Directive MDD 93/42/EEC as amended.	
	CE Certificate Number:	252.602	
	Notified Body:	NSAI National Standards Authority of Ireland.	
	MDD Device Classification:	Class IIa	